

RPFT^{Q&As}

Registry Examination for Advanced Pulmonary Function Technologists

Pass Test Prep RPFT Exam with 100% Guarantee

Free Download Real Questions & Answers **PDF** and **VCE** file from:

<https://www.pass2lead.com/rpft.html>

100% Passing Guarantee
100% Money Back Assurance

Following Questions and Answers are all new published by Test Prep
Official Exam Center

- ⚙️ **Instant Download** After Purchase
- ⚙️ **100% Money Back** Guarantee
- ⚙️ **365 Days** Free Update
- ⚙️ **800,000+** Satisfied Customers



QUESTION 1

In which of the following patients would an end-tidal CO₂ most likely produce values that are inconsistent with PaCO₂?

- A. 50 year old with mild scleroderma
- B. 42 year old with moderate COPD
- C. 20 year old who smokes cigarettes
- D. 12 year old with kyphoscoliosis

Correct Answer: B

QUESTION 2

When performing quality control in a body plethysmograph using a 5-L isothermal bottle, the VTG at shutter closure are as follows:

Trial	1	2	3	4	5
V_{TG} (L)	4.91	5.09	5.04	4.86	5.01

A pulmonary function technologist should

- A. Service the mouth pressure transducer.
- B. Recalibrate the box pressure transducer.
- C. Check biological control before beginning testing.
- D. Proceed with patient testing.

Correct Answer: A

QUESTION 3

During a cardiopulmonary stress test using breath-by-breath gas analysis, a pulmonary function technologist notices that the VO₂ suddenly decreases. Which of the following may explain this change?

- 1.
The patient has achieved anaerobic threshold.
- 2.
The measurement of the expired gas volumes is inaccurate.
- 3.

O2 analyzer "phase delay" has increased.

4.

There is a leak in the tubing or mouthpiece.

A. 1, 3, and 4 only

B. 1, 2, and 3 only

C. 1, 2, and 4 only

D. 2, 3, and 4 only

Correct Answer: A

QUESTION 4

A patient is performing a flow-volume loop. A pulmonary function technologist observes equal decreases in expiratory and inspiratory flows resulting in plateaus. Patient effort is satisfactory, and he follows instructions well. The technologist should expect the patient to have a

A. Small airways obstruction

B. Restrictive disorder

C. Variable large airways obstruction

D. Fixed large airways obstruction.

Correct Answer: B

QUESTION 5

Which of the following is an appropriate reason to perform a multiple-breath nitrogen washout test?

A. Measure anatomical dead space.

B. Differentiate obstruction from restriction.

C. Detect early small airway disease.

D. Measure oxygen consumption.

Correct Answer: C

[RPFT VCE Dumps](#)

[RPFT Practice Test](#)

[RPFT Braindumps](#)